

EVALUATION OF GENOMERA® CDX SYSTEM FOR INFLUENZA AND RSV INFECTIONS



Emeline Choquet¹, Céline Chessa^{1,2}, Andy Larivière², Agnès Beby-Defaux^{1,2},
Nicolas Lévêque^{1,2}, Maxime Pichon^{1,2,*}

1. CHU La Milétrie, Infectious Agents Department, Poitiers, France
2. Poitiers University, LITEC EA4331, Faculty of Medicine & Pharmacy, Poitiers, France

* Presenting author: maxime.pichon@chu-poitiers.fr

Poster P111

BACKGROUND

EPIDEMIOLOGY OF RESPIRATORY VIRUSES

- Influenza A (FluA), influenza B (FluB) and respiratory syncytial viruses (RSV) are viral pathogens, transmitted by aerosols. These viruses cause respiratory diseases in humans of all ages, and are responsible for burden worldwide¹.
- These viruses commonly have overlapping clinical symptoms which complicates the differentiation between these infections².
- Rapid immunoassays and antigen tests have poor sensitivity. Sensitive molecular methods are required for reliable diagnosis^{3,4}.

GENOMERA CDX™ SYSTEM

- GenomEra® CDX is a molecular diagnostics platform consisting in an integrated thermal cycler and a time-resolved fluorometer⁵.
- The Instrument is used to run analyte-specific, ready-to-use GenomEra® Test Chips. Each chip contains a sample processing control (SPC) which is included in the sample preparation tube.
- The assay sequence lasts approximately 70 minutes and ends with a report of the results.

OBJECTIVES

Evaluation of the GenomEra® CDX system performances using the GenomEra FluA/B+RSV assay (FluA/B, or RSV-A/-B) on clinical respiratory specimens

METHODS

Clinical respiratory samples

- 114 nasal swabs (NS)
- 131 nasal -throat swabs (NTS)
- 54 nasopharyngeal aspirates (NPA)



Genomera® CDX System

Thermal lysis – RT-qPCR (FluA, FluB, RSV-A, and RSV-B)

Reference Method

Automatic extraction (Nimbus®) RT-qPCR Seegene Allplex® Respiratory panel 1 on CFX96® (BioRad) (FluA, FluB, RSV-A and RSV-B)

For discrepant results (BioMérieux, France)

- Automatic Extraction on Easymag®
- RT-qPCR Influenza A/B and RSV R-Gene® kits

RESULTS

Clinical performances study

Viral Target	True positive	False positive	True negative	False negative	Sensitivity % (CI95)	Specificity % (CI95)	PPV % (CI95)	NPV % (CI 95)	Accuracy % (95% CI)
FluA (n=300*)	63	1	234	1	98.4 (91.6-100)	99.6 (97.7-100)	98.4 (89.9-99.8)	99.6 (97.1-99.9)	99.3 (97.6-99.9)
FluB (n=300**)	0	1	299	0	-	99.7 (98.2-100)	-	100	99.7 (98.2-100)
RSV-A (n=300***)	27	0	270	1	96.4 (81.7-99.9)	100 (98.6-100)	100	99.6 (97.5-100)	99.7 (98.1-100)
RSV-B (n=300***)	42	1	252	1	97.7 (87.7-99.9)	99.6 (97.8-100)	97.7 (85.6-99.7)	99.6 (97.3-99.9)	99.3 (97.6-99.9)

Figure 1. Evaluation of clinical performances of GenomEra FluA/B+RSV assay. PPV : positive predictive value. NPV: negative predictive value. For FluA, one sample out of the 300 tested samples remained inconclusive after verification(*). For FluB, one sample out of the 300 tested showed FluB borderline result, even after verification (**). For RSV, two samples out of the 300 tested showed inconclusive results for RSV-A and four for RSV-B (***)

Virus	Strains	Ct Concentration (Seegene)	200-400 X Concentration				20-40 X Concentration				2-4 X Concentration				0.2-0.4 X Concentration				0.02-0.04 X Concentration				0.002-0.004 X Concentration			
			Flu A	Flu B	RSV-A	RSV-B	Flu A	Flu B	RSV-A	RSV-B	Flu A	Flu B	RSV-A	RSV-B	Flu A	Flu B	RSV-A	RSV-B	Flu A	Flu B	RSV-A	RSV-B	Flu A	Flu B	RSV-A	RSV-B
Flu B	B/Malaysia/2506/2004 (Victoria)	34,5	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG
	B/Malaysia/2506/2004 (Victoria)		POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG
	B/Wisconsin/1/2013		POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG
	B/Brisbane/60/2008		POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG
	B/Florida/4/2006		POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG
Flu A H1	A/H1N1 - A/Solomon Islands/3/2006	37,5	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG
	A/H1N1 - A/New Caledonia/20/99		POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG
	A/H1N1 - A/California/7/2009		POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG
Flu A H3	A/H3N2 - A/Hong Kong/4801/2014	37,5	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG
	A/H3N2 - A/Victoria/210/2009		POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG
	A/H3N2 - A/Brisbane/10/2007		POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG
RSV-A	Sample 1	25	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG
	Sample 2		POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG
	Sample 3		POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG
	Sample 4		POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG
	Sample 5		POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG
RSV-B	Sample 1	24,5	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG
	Sample 2		POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG
	Sample 3		POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG
	Sample 4		POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG
	Sample 5		POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG	POS	NEG	NEG	NEG

Inclusivity study.

Figure 2. Determination of the LoD in clinical specimens.

Perfect detection of the target viral strain was found for a Ct value of:
i/ 34.6 (33.3-33.5) for FluA H₁N₁pdm09
ii/ 34.9 (34.4 - 35.4) for FluA H₃N₂
iii/ 34.0 (32.3-34.9) for FluB
iv/ 31.4 (29.2-33.8) for RSV-A
v/ 28.8 (27.3-30.2) for RSV-B

➤ According to Amplirun RNA concentrations, LoD values (indicated with a red bar) corresponded respectively to: 0.1 copies/μl (FluAH1) ; 10 copies/μl (FluAH3); 100 copies/μl (RSV-A) and 1,000 copies/μl (FluB and RSV-B)

Microorganism	Sample	Ct/C°	FluA	FluB	RSV A	RSV B
<i>Mycoplasma pneumoniae</i>	Clinical	26,8	NEG	NEG	NEG	NEG
<i>Corynebacterium jeikeum</i>	Culture	1 McF	NEG	NEG	NEG	NEG
<i>Escherichia coli</i>	Culture	1 McF	NEG	NEG	NEG	NEG
<i>Haemophilus influenzae</i>	Culture	1 McF	NEG	NEG	NEG	NEG
<i>Lactobacillus sp.</i>	Culture	1 McF	NEG	NEG	NEG	NEG
<i>Moraxella catarrhalis</i>	Culture	1 McF	NEG	NEG	NEG	NEG
<i>Pseudomonas aeruginosa</i>	Culture	1 McF	NEG	NEG	NEG	NEG
<i>Staphylococcus aureus</i>	Culture	1 McF	NEG	NEG	NEG	NEG
<i>Staphylococcus epidermidis</i>	Culture	1 McF	NEG	NEG	NEG	NEG
<i>Streptococcus pneumoniae</i>	Culture	1 McF	NEG	NEG	NEG	NEG
<i>Streptococcus pyogenes</i>	Culture	1 McF	NEG	NEG	NEG	NEG
<i>Mycobacterium tuberculosis</i>	Extract	/	NEG	NEG	NEG	NEG
<i>Chlamydia pneumoniae</i>	Clinical	32,3	NEG	NEG	NEG	NEG

Cross-reactivity study

Figure 3. Determination of the analytical specificity with other pathogens. Ct/C°: Ct value or microorganism concentration (McF=MacFarland)

Microorganism	Sample	Ct/C°	Flu A	Flu B	RSV A	RSV B
Rhinovirus	Clinical	21,2	NEG	NEG	NEG	NEG
Coronavirus OC43	Clinical	16,4	NEG	NEG	NEG	NEG
Coronavirus 229E	Clinical	19,3	NEG	NEG	NEG	NEG
Métapneumovirus	Clinical	20,5	NEG	NEG	NEG	NEG
Parainfluenzae 3	Clinical	20,2	NEG	NEG	NEG	NEG
Entérovirus	Clinical	24,5	NEG	NEG	NEG	NEG
Adénovirus	Clinical	17,2	NEG	NEG	NEG	NEG
Parainfluenzae 1	Clinical	26,1	NEG	NEG	NEG	NEG

- Cross reactivity was evaluated on highly concentrated culture (bacterial strain) or clinical samples (viral strains),
- Fourteen bacteria were tested (incl. *Mycobacterium tuberculosis* after lysis).
- Eight viral strains were tested in clinical samples with high Ct values (16.4 to 26.0).
- **No cross-reactivity could be observed on duplicates for all viral targets.**

INTERPRETATION AND CONCLUSION

A very easy-to-use, fast and powerful technology that could be applied to clinical complicated situations

- With 1 false negative and 5 false positive (including one IBV) results, performances for Influenza detection are very good but remain perfectible
 - Modification of the threshold to optimize results : no False positive results
 - **Perfect specificity and almost perfect sensitivity (Accuracy : 99.3% and 99.7% for IAV and IBV respectively)**
- With 1 false negative result, performances for RSV detection are comparable to our reference method :
 - near perfect **specificity and sensitivity (Accuracy : 99.7% and 99.3% for RSV-A and RSV-B respectively)**

A multicenter study has to be performed to validate comparison with different analytical platforms

- **Work-in-progress** (collaboration between CHU Poitiers, France; Pärnu Hospital, Estonia; FIMLAB, Jyväskylä, Finland; Vaasa Central Hospital, Finland; Turku University Hospital, Finland)

REFERENCES

1. Hodinka R.L. Respiratory RNA Viruses. Microbiol Spectr. 2016 Aug;4(4).
2. Faisey A.R., Walsh E.E. Respiratory Syncytial Virus Infection in Adults. Clin Microbiol Rev. 2000 Jul; 13.
3. "Guidance for Clinicians on the Use of Rapid Influenza Diagnostic Tests". Centers for Disease Control and Prevention. Jan 18, 2017. www.cdc.gov/flu/professionals/diagnosis/clinician_guidance_ridt.htm.
4. Chartrand C., Tremblay N., Renaud C., Pappenburg J. Diagnostic Accuracy of Rapid Antigen Detection Tests for Respiratory Syncytial Virus Infection: Systematic Review and Meta-analysis. J Clin Microbiol. 2015 Dec;53(12):3738-49.
5. Hagren V., von Lode P., Syyräjä A., Soukka T., Lövgren T., Kojoja H., Nurmi J. 2008. An automated PCR platform with homogeneous time-resolved fluorescence detection and dry chemistry assay kits. Anal Biochem. 374:411-6.